



# ***Dräger CMS Validation***

Sarah Anderson  
Bowie High School  
8/6/04

Naval Research Laboratory-Chemistry  
Division


Mentors: Dr. S. Rose-Pehrsson &  
Kimberly Parker

**S  
E  
A  
P  
.  
O  
4**

## Introduction

- Submarine Atmosphere- Unique environment- Sailor extended exposure.
- Hard to detect- exposure to hazardous chemical and biological agents.
- Sailors on Navy submarines need equipment to detect these easily and fast.
- **Dräger** tubes being used currently.
  - Manually operated pumps
  - Human error



- 
- Evaluate ***Dräger CMS*** response
    - To known concentrations of applied gasses
    - To the interference gas (Hydrogen, the submarines normal background) introduced.
    - Two tests for applied gasses, five trials each.
  - Hydrogen(H<sub>2</sub>) Background tested
    - 300ppm
    - Submarine background varies from 200ppm-10,000ppm


- ***Dräger CMS (Chip Measurement System)-  
an electronic based analyzer with integrated  
Data Recorder and  
remote system.***



- ***Chips-Substance-  
specific colorimetric  
measuring chips:***




# Vapor Generation System



# *Dräger CMS Results*

## Data/results

% Relative Standard Deviation



# Table 1

Tests Calculations	Tests Without H <sub>2</sub>	Tests With H <sub>2</sub>	Tests Without H <sub>2</sub>	Tests With H <sub>2</sub>	Tests Without H <sub>2</sub>	Tests With H <sub>2</sub>	Tests Without H <sub>2</sub>	Tests With H <sub>2</sub>
Chemical name	<i>Nitrogen Dioxide</i>	<i>Nitrogen Dioxide</i>	<i>Benzene</i>	<i>Benzene</i>	<i>Hydrogen Sulfide</i>	<i>Hydrogen Sulfide</i>	<i>Sulphur Dioxide</i>	<i>Sulphur Dioxide</i>
Concentration Tested	6ppm	6ppm	1ppm	1ppm	10ppm	10ppm	15ppm	15ppm
Standard Deviation	0.36ppm	0.19ppm	0.68ppm	0.82ppm	0.51ppm	0.68ppm	1.13ppm	1.08ppm
%RSD	6.22%	3.27%	56.49%	71.47%	4.78%	6.77%	9.76%	9.13%
Student's T-Test	95% chance No significant Difference		95% chance No significant Difference		95% chance No significant Difference		95% chance No significant Difference	


# Table 2

Tests Calculations	Tests Without H2	Tests With H2	Tests Without H2	Tests With H2
Chemical name	<i>Carbon Monoxide</i>	<i>Carbon Monoxide</i>	<i>Ammonia</i>	<i>Ammonia</i>
Concentration Tested	50ppm	50ppm	10ppm	10ppm
Standard Deviation	4.28ppm	2.30ppm	0.50ppm	0.47ppm
%RSD	10.54%	6.50%	5.33%	4.53%
Student's T-Test	95% chance there IS a significant difference		95% chance there IS a significant difference	




# Hydrogen Interference

## Carbon Monoxide



## Ammonia



# Conclusion



- Stable and Reproducible results
- %RSD <10%
  - Except for the pump failure on Benzene
- Suggested Hydrogen interference in Ammonia and Carbon Monoxide tests.
- Overall, reliable in testing known analytes in known concentrations.

## ■ Future

- a universal detector vs. individual sample for one analyte.

## ■ Acknowledgments

- I would like to thank
  - SEAP Program
  - NRL-Chemistry Division
  - My mentors Dr. S. Rose-Pehrsson &  
Kimberly Parker & Mark Hammond



## ■ THANK YOU